VAHTS Maxi Unique Dual Index Primers Set 2 for Illumina

N34402

Version 24.2



Product Description

VAHTS Maxi Unique Dual Index Primers Set 2 for Illumina is a specialized kit for DNA library preparation based on the Illumina sequencers. The VAHTS Universal Adapter for Illumina in the kit is the stubby adapter that can be compatible with various library preparation kits based on the T-A ligation. VAHTS Unique Dual Index Primer for Illumina contains specifically paired Unique Dual Index (UDI). After library preparation, it can achieve two distinct and unrelated index sequences double verification, minimizing index hopping and index misassignment, ensuring that the reads in downstream analysis are nearest to the truth.

This kit includes Universal Adapters and 96 different VAHTS Unique Dual Index Primer for Illumina. All the reagents provided in the kit have undergone rigorous quality control and functional testing to ensure the optimal stability and repeatability of library preparation.

Components

Components	N34402-01 (384 rxns)
VAHTS Unique Dual Index Primer for Illumina	20 μl each
VAHTS Universal Adapter for Illumina	4 × 480 μl

[▲] The amount of DNA Adapter used for a DNA library depends on the amount of input DNA. For details, please refer to the Appendix.

N34402 Components

	1	2	3	4	5	6	7	8	9	10
1	097	107	387	127	390	147	157	167	177	187
2	098	108	118	128	138	148	158	168	178	188
3	099	109	119	129	139	149	159	169	179	189
4	100	110	120	130	391	150	160	170	180	190
5	101	111	121	131	141	151	161	171	181	191
6	386	112	388	132	392	152	162	172	182	192
7	103	113	123	133	143	153	163	173	393	Ada
8	104	114	124	134	144	154	164	174	184	Ada
9	105	115	125	389	145	155	165	175	185	Ada
10	106	116	126	136	146	156	166	176	186	Ada

[▲] The 12 Indexes marked with the same color in the table can ensure base balance and are recommended to be used in groups.

Storage

Store at -30 ~ -15°C and ship at ≤0°C.

Applications

VAHTS Maxi Unique Dual Index Primers Set 2 for Illumina (Vazyme #N34402) is a specialized kit for Vazyme ND/NR series library preparation kits, compatible with various sample types: genomic DNA, cfDNA, FFPE DNA, etc. It is suitable for mutiple dual-index DNA libraries for Illumina sequencers, and can effectively avoid cross-contamination between samples.

Notes

- 1. The amount of DNA Adapter used depends on the amount of input DNA. For details, please refer to the Appendix.
- 2. Don't premix Adapter, Ligation Buffer and Ligase to prevent adapter self-ligation.
- 3. Don't store the adapters above room temperature (15 ~ 25°C) which may result in decreasing ligation efficiency.

A The concentration of VAHTS Universal Adapter for Illumina is 15 μM; the concentration of VAHTS Unique Dual Index Primer for Illumina is 12 μΜ.

Appendix

Table 1. Recommended the Dilution Ratio and Volume of Adapters for Vazyme #ND610 (Mechanical Fragmentation)

Input DNA	Adapter dilution ratio	Volume (µI)
0.1 ng	1:60	5
1 ng	1:10	5
10 ng	1:2	5
50 ng	Undiluted	5
≥100 ng	Undiluted	5

Table 2. Recommended the Dilution Ratio and Volume of Adapters for Vazyme #ND627 (Enzymatic Fragmentation)

Input DNA	Adapter dilution ratio	Volume (µI)
1 ng	1:10	5
10 ng	1:2	5
50 ng	Undiluted	5
≥100 ng	Undiluted	5

Table 3. Recommended the Dilution Ratio and Volume of Adapters for Vazyme #NR605 (RNA Library Preparation)

Input RNA	Adapter dilution ratio	Volume (µI)
10 ng	1:10	0.5
100 ng	1:10	1
1 μg	1:10	3.5

For Research Use Only. Not for use in diagnostic procedures.