HiScript III U⁺ One Step qRT-PCR Probe 5 × Master Mix

Q611

Version 24.1



Product Description

HiScript III U $^+$ One Step qRT-PCR Probe 5 × Master Mix is single-tube RT-qPCR premix that is suitable for singleplex or multiplex qPCR detection using RNA (e.g. RNA virus) as template, with extremely high stability. The reaction can be performed directly after adding templates, primers and probes. Using gene specific primers (GSP), the reverse transcription and qPCR can be finished in one tube, significantly reducing pipetting procedures and the risk of contamination.

This kit can prevent the contamination of PCR product. Combining the superior performance of HiScript III Reverse Transcriptase and hot-start Champagne Taq DNA Polymerase, with an optimized buffering system, the detection sensitivity of HiScript III U⁺ One Step qRT-PCR Probe 5 × Master Mix can reach 0.1 pg of total RNA or less than 10 copies of RNA templates and is suitable for high-specificity detection systems based on fluorescence labelled probes (e.g. TaqMan). In addition, the dUTP/UDG anti-contamination system is introduced in it, which can work at room temperature to eliminate the influence of amplification product contamination on qPCR and ensure the accuracy of results.

Components

Components	Q611-01 100 rxns (20 μl/rxn)	Q611-02 1,000 rxns (20 μl/rxn)	Q611-03 10,000 rxns (20 μl/rxn)
RNase-free ddH₂O	1 ml	10 ml	100 ml
U⁺ One Step qRT-PCR Probe 5 × Master Mixª	400 µl	4 × 1 ml	40 ml
50 × ROX Reference Dye 1 ^b	40 µl	400 µl	4 × 1 ml
50 × ROX Reference Dye 2 ^b	40 µl	400 µl	4 × 1 ml

a. It contains dNTP/dUTP Mix, Mg²⁺, HiScript III Reverse Transcriptase, RNase inhibitor, Heat-labile UDG, and Champagne Taq DNA Polymerase.

Storage

Store at -30 ~ -15°C and ship at ≤0°C.

Applications

It is applicable for detection of various RNA nucleic acids of animals, plants and microorganisms (viruses, etc.).

Notes

- 1. U⁺ One Step qRT-PCR Probe 5 × Master Mix contains high concentration of glycerol. Please centrifuge briefly and mix gently before use.
- 2. To avoid contamination, please use RNase-free tips and EP tubes.

b. It is used to correct the error of fluorescence signals between wells. Use 50 × ROX Reference Dye 1 for ABI 7900HT/7300 Real-Time PCR System and StepOnePlus; Use 50 × ROX Reference Dye 2 for ABI 7500, 7500 Fast Real-Time PCR System, and Stratagene Mx3000P. Don't use ROX for Roche and Bio-Rad Real-Time PCR instruments.

Experiment Process (Using ABI StepOnePlus as a test machine)

1. Prepare the reaction solution in an RNase-free PCR tube

Components	Volume	
RNase-free ddH ₂ O	to 20 μl	
U⁺ One Step qRT-PCR Probe 5 × Master Mix	4 μΙ	
50 × ROX Reference Dye 1	0.4 μΙ	
Primer Forward (10 μM)	0.4 μΙ	
Primer Reverse (10 µM)	0.4 μΙ	
Taqman Probe (10 μM)	0.2 μΙ	
Template RNA	Total RNA: 1 pg - 1 μg	

For each component, the recommended volume can be adjusted as follows:

Generally, the final concentration of primer should be 0.2 μ M. If necessary, it can be adjusted in the range of 0.1 - 1.0 μ M.

- ▲ The final concentration of TaqMan Probe can be adjusted between 50 250 nM.
- ▲ The accuracy of template volumes have significant impacts on the qPCR results, due to the high sensitivity of qPCR. Therefore, to improve the experimental repeatability, it is recommended to dilute the template and pipet more volume (e.g. diluted to 2 5 µl/sample) to the reaction system.
- ▲ The size of the amplified products should be within the range of 80 200 bp.

2. Reaction Program

Standard Program (for the optimal amplification sensitivity)

Stage 1	Reverse Transcription	Rep: 1	55°Cª	15 min
Stage 2	Initial Denaturation	Rep: 1	95°C	30 sec
Stage 3	Cycles	Reps: 45	95°C	10 sec
			60°C	30 sec
t Program (su	itable for most One Step gRT-PCR	applications)		
	itable for most One Step qRT-PCR Reverse Transcription	applications) Rep: 1	55°Cª	5 min
Stage 1			55°C° 95°C	
st Program (su Stage 1 Stage 2 Stage 3	Reverse Transcription	Rep: 1		5 min 30 sec 5 sec

a. For templates with complex secondary structure or high GC content, the reverse transcription temperature can be increased to 55°C, which will improve the sensitivity and performance.

3. Data analysis of the Real Time PCR amplification curve and the standard curve, etc.

b. The extension time varies between different qPCR instruments used. For ABI 7700 and 7900HT, the extension time should be ≥30 sec; for ABI 7000 and 7300, the extension time should be ≥31 sec; and for ABI 7500, ≥34 sec;

c. Please check if the fast program is compatible with the qPCR instrument.